

Manufacturer Disclosure Statement for Medical Device Security – MDS <sup>2</sup>				
DEVICE DESCRIPTION				
Device Category	Manufacturer	Document ID	Document Release Date	
16247	Carestream Health, Inc.	7K7078	04/16/2016	
Device Model	Software Revision	Software Release Date		
CARESTREAM Image Suite	4.0 and Higher	12/12/2013		
Manufacturer or Representative Contact Information	Company Name	Manufacturer Contact Information		
	Technical Support	Telephone #: 1-800-328-2910	e-mail: health.imaging.tsc@carestreamhealth.com	
Intended use of device in network-connected environment:				
MANAGEMENT OF PRIVATE DATA				
Refer to Section 2.3.2 of this standard for the proper interpretation of information requested in this form.			Yes, No, N/A, or See Note	Note #
A	Can this <b>device</b> display, transmit, or maintain <b>private data</b> (including <b>electronic Protected Health Information</b> [ePHI])?		Yes	—
B	Types of <b>private data</b> elements that can be maintained by the <b>device</b> :			
	B.1	Demographic (e.g., name, address, location, unique identification number)?	Yes	—
	B.2	Medical record (e.g., medical record #, account #, test or treatment date, <b>device</b> identification number)?	Yes	—
	B.3	Diagnostic/therapeutic (e.g., photo/radiograph, test results, or physiologic data with identifying characteristics)?	Yes	—
	B.4	Open, unstructured text entered by <b>device user/operator</b> ?	No	—
	B.5	<b>Biometric data</b> ?	No	—
	B.6	Personal financial information?	No	—
C	Maintaining <b>private data</b> - Can the <b>device</b> :			
	C.1	Maintain <b>private data</b> temporarily in volatile memory (i.e., until cleared by power-off or reset)?	Yes	—
	C.2	Store <b>private data</b> persistently on local media?	Yes	—
	C.3	Import/export <b>private data</b> with other systems?	Yes	—
	C.4	Maintain <b>private data</b> during power service interruptions?	Yes	1
D	Mechanisms used for the transmitting, importing/exporting of <b>private data</b> – Can the <b>device</b> :			
	D.1	Display private data (e.g., video display, etc.)?	Yes	—
	D.2	Generate hardcopy reports or images containing <b>private data</b> ?	Yes	—
	D.3	Retrieve <b>private data</b> from or record <b>private data</b> to <b>removable media</b> (e.g., disk, DVD, CD-ROM, tape, CF/SD card, memory stick, etc.)?	Yes	—
	D.4	Transmit/receive or import/export <b>private data</b> via dedicated cable connection (e.g., IEEE 1073, serial port, USB, FireWire, etc.)?	Yes	—
	D.5	Transmit/receive <b>private data</b> via a wired network connection (e.g., LAN, WAN, VPN, intranet, Internet, etc.)?	Yes	—
	D.6	Transmit/receive <b>private data</b> via an integrated wireless network connection (e.g., WiFi, Bluetooth, infrared, etc.)?	Yes	—
	D.7	Import <b>private data</b> via scanning?	Yes	—
	D.8	Other?	No	—
Management of Private Data notes:			1. All data is written to the system hard drive. To protect against data loss/corruption it is strongly recommended to install UPS equipment to provide uninterruptible power.	

© Copyright 2013 by the National Electrical Manufacturers Association and the Healthcare Information and Management Systems Society.

Device Category 16247	Manufacturer Carestream Health, Inc.	Document ID 7K7078	Document Release Date 04/16/2016
Device Model CARESTREAM Image Suite	Software Revision 4.0 and Higher		Software Release Date 12/12/2013

**SECURITY CAPABILITIES**

Refer to Section 2.3.2 of this standard for the proper interpretation of information requested in this form.		Yes, No, N/A, or See Note	# Note
<b>1</b>	<b>AUTOMATIC LOGOFF (ALOF)</b> The <b>device's</b> ability to prevent access and misuse by unauthorized <b>users</b> if <b>device</b> is left idle for a period of time.		
1-1	Can the <b>device</b> be configured to force reauthorization of logged-in <b>user(s)</b> after a predetermined length of inactivity (e.g., auto-logoff, session lock, password protected screen saver)?	Yes	—
1-1.1	Is the length of inactivity time before auto-logoff/screen lock <b>user</b> or administrator configurable? (Indicate time [fixed or configurable range] in notes.)	Yes	—
1-1.2	Can auto-logoff/screen lock be manually invoked (e.g., via a shortcut key or proximity sensor, etc.) by the <b>user</b> ?	Yes	—
ALOF notes:			
<b>2</b>	<b>AUDIT CONTROLS (AUDT)</b> The ability to reliably audit activity on the <b>device</b> .		
2-1	Can the <b>medical device</b> create an <b>audit trail</b> ?	Yes	—
2-2	Indicate which of the following events are recorded in the audit log:		
2-2.1	Login/logout	Yes	—
2-2.2	Display/presentation of data	Yes	—
2-2.3	Creation/modification/deletion of data	Yes	—
2-2.4	Import/export of data from <b>removable media</b>	Yes	—
2-2.5	Receipt/transmission of data from/to external (e.g., network) connection	Yes	—
2-2.5.1	<b>Remote service</b> activity	Yes	2
2-2.6	Other events? (describe in the notes section)	No	—
2-3	Indicate what information is used to identify individual events recorded in the audit log:		
2-3.1	<b>User ID</b>	See Note	3
2-3.2	Date/time	Yes	—
AUDT notes: 2. System has internal login control and activity logging. 3. Some Logging events contain the User ID, others do not.			
<b>3</b>	<b>AUTHORIZATION (AUTH)</b> The ability of the device to determine the authorization of users.		
3-1	Can the <b>device</b> prevent access to unauthorized <b>users</b> through <b>user</b> login requirements or other mechanism?	Yes	—
3-2	Can <b>users</b> be assigned different privilege levels within an application based on 'roles' (e.g., guests, regular <b>users</b> , power <b>users</b> , administrators, etc.)?	Yes	—
3-3	Can the <b>device</b> owner/ <b>operator</b> obtain unrestricted administrative privileges (e.g., access operating system or application via local root or admin account)?	No	—
AUTH notes:			

Device Category 16247	Manufacturer Carestream Health, Inc.	Document ID 7K7078	Document Release Date 04/16/2016	
Device Model CARESTREAM Image Suite	Software Revision 4.0 and Higher	Software Release Date 12/12/2013		
Refer to Section 2.3.2 of this standard for the proper interpretation of information requested in this form.			Yes, No, N/A, or See Note	# Note
<b>4 CONFIGURATION OF SECURITY FEATURES (CNFS)</b>				
The ability to configure/re-configure <b>device security capabilities</b> to meet <b>users'</b> needs.				
4-1	Can the <b>device</b> owner/operator reconfigure product <b>security capabilities</b> ?		Yes	4
4. The Web Site can configured to access by Http or Https.				
CNFS notes:				
<b>5 CYBER SECURITY PRODUCT UPGRADES (CSUP)</b>				
The ability of on-site service staff, remote service staff, or authorized customer staff to install/upgrade <b>device's</b> security patches.				
5-1	Can relevant OS and <b>device</b> security patches be applied to the <b>device</b> as they become available?		Yes	—
	5-1.1	Can security patches or other software be installed remotely?	Yes	5
5. Through installed remote access tool, service is able to install patches and other software remotely.				
CSUP notes:				
<b>6 HEALTH DATA DE-IDENTIFICATION (DIDT)</b>				
The ability of the <b>device</b> to directly remove information that allows identification of a person.				
6-1	Does the <b>device</b> provide an integral capability to de-identify <b>private data</b> ?		Yes	6
6. When export image as Patient CD, user can select the option to hide patient information (Patient Name/Patient ID, etc.)				
DIDT notes:				
<b>7 DATA BACKUP AND DISASTER RECOVERY (DTBK)</b>				
The ability to recover after damage or destruction of <b>device</b> data, hardware, or software.				
7-1	Does the <b>device</b> have an integral data backup capability (i.e., backup to remote storage or <b>removable media</b> such as tape, disk)?		Yes	7
7. This could be accomplished by the Windows operating systems backup / restore capability. For Patient data, the system may burn Patient CD and archive image to CD/DVD or removable USB disk.				
DTBK notes:				
<b>8 EMERGENCY ACCESS (EMRG)</b>				
The ability of <b>device users</b> to access <b>private data</b> in case of an emergency situation that requires immediate access to stored <b>private data</b> .				
8-1	Does the <b>device</b> incorporate an <b>emergency access</b> ("break-glass") feature?		Yes	—
EMRG notes:				
<b>9 HEALTH DATA INTEGRITY AND AUTHENTICITY (IGAU)</b>				
How the <b>device</b> ensures that data processed by the <b>device</b> has not been altered or destroyed in an unauthorized manner and is from the originator.				
9-1	Does the <b>device</b> ensure the integrity of stored data with implicit or explicit error detection/correction technology?		Yes	
IGAU notes:				

© Copyright 2013 by the National Electrical Manufacturers Association and the Healthcare Information and Management Systems Society.

Device Category 16247	Manufacturer Carestream Health, Inc.	Document ID 7K7078	Document Release Date 04/16/2016	
Device Model CARESTREAM Image Suite	Software Revision 4.0 and Higher	Software Release Date 12/12/2013		
Refer to Section 2.3.2 of this standard for the proper interpretation of information requested in this form.			Yes, No, N/A, or See Note	Note #
<b>10 MALWARE DETECTION/PROTECTION (MLDP)</b>				
The ability of the <b>device</b> to effectively prevent, detect and remove malicious software ( <b>malware</b> ).				
10-1	Does the <b>device</b> support the use of <b>anti-malware</b> software (or other <b>anti-malware</b> mechanism)?		No	—
10-1.1	Can the <b>user</b> independently re-configure <b>anti-malware</b> settings?		No	—
10-1.2	Does notification of <b>malware</b> detection occur in the <b>device user</b> interface?		No	—
10-1.3	Can only manufacturer-authorized persons repair systems when <b>malware</b> has been detected?		No	—
10-2	Can the device owner install or update <b>anti-virus software</b> ?		Yes	8
10-3	Can the device owner/ <b>operator</b> (technically/physically) update virus definitions on manufacturer-installed <b>anti-virus software</b> ?		Yes	8
MLDP notes:	8. It is strongly recommended to install anti-virus/firewall software at both the host PC and web client PC, McAfee software is recommended to be installed along with Image Suite application.			
<b>11 NODE AUTHENTICATION (NAUT)</b>				
The ability of the <b>device</b> to authenticate communication partners/nodes.				
11-1	Does the <b>device</b> provide/support any means of node authentication that assures both the sender and the recipient of data are known to each other and are authorized to receive transferred information?		Yes	—
NAUT notes:				
<b>12 PERSON AUTHENTICATION (PAUT)</b>				
Ability of the <b>device</b> to authenticate <b>users</b>				
12-1	Does the <b>device</b> support <b>user/operator</b> -specific username(s) and password(s) for at least one <b>user</b> ?		Yes	—
12-1.1	Does the device support unique <b>user/operator</b> -specific IDs and passwords for multiple users?		Yes	—
12-2	Can the <b>device</b> be configured to authenticate <b>users</b> through an external authentication service (e.g., MS Active Directory, NDS, LDAP, etc.)?		No	—
12-3	Can the <b>device</b> be configured to lock out a <b>user</b> after a certain number of unsuccessful logon attempts?		No	—
12-4	Can default passwords be changed at/prior to installation?		No	—
12-5	Are any shared <b>user</b> IDs used in this system?		Yes	—
12-6	Can the <b>device</b> be configured to enforce creation of <b>user</b> account passwords that meet established complexity rules?		Yes	—
12-7	Can the <b>device</b> be configured so that account passwords expire periodically?		Yes	—
PAUT notes:				
<b>13 PHYSICAL LOCKS (PLOK)</b>				
Physical locks can prevent unauthorized <b>users</b> with physical access to the <b>device</b> from compromising the integrity and confidentiality of <b>private data</b> stored on the <b>device</b> or on <b>removable media</b> .				
13-1	Are all <b>device</b> components maintaining <b>private data</b> (other than <b>removable media</b> ) physically secure (i.e., cannot remove without tools)?		No	9
PLOK notes:	9. It is the site's responsibility to protect the PC equipment. (For example, by locking the PC to the table etc.)			

Device Category 16247	Manufacturer Carestream Health, Inc.	Document ID 7K7078	Document Release Date 04/16/2016	
Device Model CARESTREAM Image Suite	Software Revision 4.0 and Higher	Software Release Date 12/12/2013		
Refer to Section 2.3.2 of this standard for the proper interpretation of information requested in this form.			Yes, No, N/A, or See Note	# Note
<b>14 ROADMAP FOR THIRD PARTY COMPONENTS IN DEVICE LIFE CYCLE (RDMP)</b>				
Manufacturer's plans for security support of 3rd party components within <b>device</b> life cycle.				
14-1	In the notes section, list the provided or required (separately purchased and/or delivered) operating system(s) - including version number(s).		Yes	10
14-2	Is a list of other third party applications provided by the manufacturer available?  10.Win 7 SP1/Win 8.1		Yes	—
RDMP notes:				
<b>15 SYSTEM AND APPLICATION HARDENING (SAHD)</b>				
The <b>device's</b> resistance to cyber attacks and <b>malware</b> .				
15-1	Does the <b>device</b> employ any hardening measures? Please indicate in the notes the level of conformance to any industry-recognized hardening standards.		No	—
15-2	Does the <b>device</b> employ any mechanism (e.g., release-specific hash key, checksums, etc.) to ensure the installed program/update is the manufacturer-authorized program or software update?		No	—
15-3	Does the <b>device</b> have external communication capability (e.g., network, modem, etc.)?		Yes	—
15-4	Does the file system allow the implementation of file-level access controls (e.g., New Technology File System (NTFS) for MS Windows platforms)?		No	—
15-5	Are all accounts which are not required for the <b>intended use</b> of the <b>device</b> disabled or deleted, for both <b>users</b> and applications?		No	—
15-6	Are all shared resources (e.g., file shares) which are not required for the <b>intended use</b> of the <b>device</b> , disabled?		No	—
15-7	Are all communication ports which are not required for the <b>intended use</b> of the <b>device</b> closed/disabled?		No	—
15-8	Are all services (e.g., telnet, file transfer protocol [FTP], internet information server [IIS], etc.), which are not required for the <b>intended use</b> of the <b>device</b> deleted/disabled?		No	—
15-9	Are all applications (COTS applications as well as OS-included applications, e.g., MS Internet Explorer, etc.) which are not required for the <b>intended use</b> of the <b>device</b> deleted/disabled?		No	—
15-10	Can the <b>device</b> boot from uncontrolled or <b>removable media</b> (i.e., a source other than an internal drive or memory component)?		No	—
15-11	Can software or hardware not authorized by the <b>device</b> manufacturer be installed on the device without the use of tools?  11. Only privileged users in operating system can reach the desktop and install additional software on the PC. The medical device is the Image Suite application software that is installed on a PC meeting the minimum specifications. It is the responsibility of the user to secure the PC and its OS.		Yes	11
SAHD notes:				
<b>16 SECURITY GUIDANCE (SGUD)</b>				
The availability of security guidance for <b>operator</b> and administrator of the system and manufacturer sales and service.				
16-1	Are security-related features documented for the <b>device user</b> ?		Yes	—
16-2	Are instructions available for <b>device/media</b> sanitization (i.e., instructions for how to achieve the permanent deletion of personal or other sensitive data)?		Yes	—
SGUD notes:				

Device Category 16247	Manufacturer Carestream Health, Inc.	Document ID 7K7078	Document Release Date 04/16/2016	
Device Model CARESTREAM Image Suite	Software Revision 4.0 and Higher	Software Release Date 12/12/2013		
Refer to Section 2.3.2 of this standard for the proper interpretation of information requested in this form.			Yes, No, N/A, or See Note	Note #
<b>17 HEALTH DATA STORAGE CONFIDENTIALITY (STCF)</b>				
The ability of the <b>device</b> to ensure unauthorized access does not compromise the integrity and confidentiality of <b>private data</b> stored on <b>device</b> or <b>removable media</b> .				
17-1	Can the <b>device</b> encrypt data at rest?		No	—
STCF notes:				
<b>18 TRANSMISSION CONFIDENTIALITY (TXCF)</b>				
The ability of the <b>device</b> to ensure the confidentiality of transmitted <b>private data</b> .				
18-1	Can <b>private data</b> be transmitted only via a point-to-point dedicated cable?		No	—
18-2	Is <b>private data</b> encrypted prior to transmission via a network or <b>removable media</b> ? (If yes, indicate in the notes which encryption standard is implemented.)		Yes	12
18-3	Is <b>private data</b> transmission restricted to a fixed list of network destinations?		Yes	13
TXCF notes: 12. Provide option to transfer PHI data via a network using HTTPS secure connection. All data is encrypted prior to transmission. But, still not encrypted prior to transmission via a removal media. 13. PHI data is sent to those predefined DICOM targets only.				
<b>19 TRANSMISSION INTEGRITY (TXIG)</b>				
The ability of the <b>device</b> to ensure the integrity of transmitted <b>private data</b> .				
19-1	Does the <b>device</b> support any mechanism intended to ensure data is not modified during transmission? (If yes, describe in the notes section how this is achieved.)		No	—
TXIG notes:				
<b>20 OTHER SECURITY CONSIDERATIONS (OTHR)</b>				
Additional security considerations/notes regarding <b>medical device</b> security.				
20-1	Can the <b>device</b> be serviced remotely?		Yes	14
20-2	Can the <b>device</b> restrict remote access to/from specified devices or <b>users</b> or network locations (e.g., specific IP addresses)?		Yes	15
20-2.1	Can the <b>device</b> be configured to require the local <b>user</b> to accept or initiate remote access?		No	—
OTHR notes: 14. Requires the site to be connected to the network and allow remote access. 15. Requires modifications to the Window's OS definitions.				